<http://www.diva-portal.org>

Preprint

This is the submitted version of a paper published in *Applied Optics*.

Citation for the original published paper (version of record):

Stangner, T., Hanqing, Z., Tobias, D., Krister, W., Andersson, M. (2017)

Step-by-step guide to reduce spatial coherence of laser light using a rotating ground glass diffuser.

Applied Optics, 56(19): 5427-5435

<https://doi.org/10.1364/AO.56.005427>

Access to the published version may require subscription.

N.B. When citing this work, cite the original published paper.

Permanent link to this version:

<http://urn.kb.se/resolve?urn=urn:nbn:se:umu:diva-135625>

A step-by-step guide to reduce spatial coherence of laser light using a rotating ground glass diffuser

TIM STANGNER¹, HANQING ZHANG¹, TOBIAS DAHLBERG¹, KRISTER WIKLUND¹, AND MAGNUS ANDERSSON^{1,*}

¹Department of Physics, Umeå University, 901 87 Umeå, Sweden

*Corresponding author: magnus.andersson@umu.se

Compiled May 30, 2017

Wide field-of-view imaging of fast processes in a microscope requires high light intensities motivating the use of lasers as light sources. However, due to their long spatial coherence length lasers are inappropriate for such applications as they produce coherent noise and parasitic reflections, such as speckle, degrading image quality. Therefore, we provide a step-by-step guide for constructing a speckle-free and high contrast laser illumination setup using a rotating ground glass diffuser driven by a stepper motor. The setup is easy to build, cheap and allows a significant light throughput of 48 %, which is 40 % higher in comparison to a single lens collector commonly used in reported setups. This is achieved by using only one objective to collect the scattered light from the ground glass diffuser. We validate our setup in terms of image quality, speckle contrast, motor-induced vibrations and light throughput. To highlight the latter, we record Brownian motion of micro-particles using a 100x oil immersion objective and a high-speed camera operating at 2000 Hz with a laser output power of only 22 mW. Moreover, by reducing the objective magnification to 50x sampling rates up to 10 000 Hz are realized. To help readers with basic or advanced optics knowledge realizing this setup we provide; a full component list, 3D-printing CAD files, setup protocol, and the code for running the stepper motor. © 2017 Optical Society of America

OCIS codes: (120.4820) Optical systems; (140.3460) Lasers; (230.1980) Diffusers; (110.0110) Imaging systems; (110.4980) Partial coherence in imaging;

<http://dx.doi.org/10.1364/ao.XX.XXXXXX>

1. INTRODUCTION

Visualizing fast processes such as particle, cell or protein diffusion using high-speed cameras, necessitates light sources with high intensity output and low spatial coherence, like random lasers [1], superluminescent diodes (SLD) [2, 3], laser driven light sources (LDLS) [4] or light emitting diodes (LED). However, images produced by these sources have limited resolution and signal-to-noise ratio (LDLS, LED); or are expensive and have restricted usability (random laser, SLD); and might not be part of standard laboratory equipment. In contrast, lasers are rather cheap and can be found in almost each laboratory and their intrinsic properties can provide high signal-to-noise ratio imaging.

Though in terms of sample illumination, their long spatial coherence length is problematic, since coherent noise such as subjective speckles introduces artifacts in microscopic images. In this context, subjective speckles refer to a random granular pattern originating from interference between highly coherent light and its back reflections from rough surfaces, such as glass [5–7]. In presences of subjective speckle pattern, image quality is

poor making it impossible to reveal detailed information about the system under study [8].

To reduce speckle contrast and spatial coherence of lasers, several methods have been suggested including, induction of vibrations in a section of an optical fiber [9], scrambling the light by focusing the laser beam into a multi-mode optical fiber [10] or guiding the laser beam on a holographic diffuser [11]. However, one of the most common approach to remove coherent noise is positioning a rotating or vibrating ground glass diffuser in the beam path, producing a partially coherent light source [12–24]. To collect the scattered light from a diffuser single lenses with different focal lengths [16–20] or beam splitter cubes [21–23] are used in literature, offering only modest light collection. Additionally, by using a single lens as collector, the width of the collimated beam is in the order of the lens diameter, typically 25.4 mm, and needs to be reduced afterwards by adding additional lenses and therefore increasing alignment effort significantly.

Even-though the approach to reduce spatial coherence using a rotating ground glass is known for almost 40 years, replicating

Fig. 1. Photograph (a) and schematic (b) of the speckle-free laser illumination setup. The laser beam is centered on a $f = 100$ mm lens (L1) by two mirrors (M1, M2), focusing the light on the rotating ground glass diffuser (RGG). Rotation of the RGG is accomplished by using a stepper motor (SM). The scattered light from RGG is collected with the collection objective (CO), producing a parallel beam output at its back aperture. Subsequently, the light is focused into a multi-mode optical fiber (OF) using a focusing objective (FO).

such a setup remains challenging since experimental details are rarely published. Especially, details about motor choice and control, how to mount the ground glass diffuser on the motor, motor rotation speed and how to ensure a high light throughput, are crucial to assemble a robust setup but also to keep construction time and cost low.

To overcome issues with motor choice and control, commercial solutions are available such as motorized rotation stages. However, these stages are made to achieve high angular precision, usually only offering rotation speeds in the range of 0.1–3 Hz which is too slow for effective speckle averaging at short camera shutter times. In addition, these precision motorized rotation stages and their respective drivers are expensive with prices up to several thousands of USD.

Therefore, we provide a simple guide to build a cheap and robust speckle-free laser illumination setup providing a high light throughput to realize camera sampling rates up to 10 000 Hz using only laser powers in the low mW-range. The centerpiece of the setup is a rotating ground glass diffuser mounted on a stepper motor and one objective to collect scattered light behind the diffuser. The total cost for stepper motor, its electrical controller unit and objective is below 500 USD. We document each construction and alignment step and characterize our setup in terms of image quality, speckle contrast, light throughput and motor-induced vibrations. To highlight a light throughput of 48 %, we investigate the Brownian motion of micro-particles using a high-speed camera. By providing speckle-free images with excellent signal-to-noise ratio, this setup can be used as a light source for digital holographic microscopy [25] and optical coherence tomography [26]. Additionally, we provide; a full component list, CAD-files for constructing mounts and holders, a code to control stepper motors and a Matlab program to calculate the speckle contrast, allowing readers to upgrade the design to fit their experimental requirements.

2. MATERIALS AND METHODS

A. Building a speckle-free laser illumination setup using a stepper motor and a rotating ground glass diffuser

The speckle-free laser illumination setup (Fig. 1 a,b) is realized by focusing a laser beam on a rotating ground glass diffuser and coupling the scattered light into an optical fiber. As light source, we use a single frequency continuous wave (CW) diode pumped laser (Cobolt 04-01 Series, Calypso™50, $\lambda = 491$ nm, max. output power 45 mW, Cobolt AB, Solna, Sweden), possessing good optical properties such as low intensity noise, high pointing stability and a coherence length of ≈ 300 m. In general, however, any laser can be used as light source since the speckle pattern depends only on the angle, polarization, and wavelength of the illuminating laser beam [27].

In the first alignment step, we couple the laser light, using a focusing objective (FO), into the multi-mode optical fiber (OF) that is mounted on a *xyz*-stage of a fiber launch system. Please note, focusing light into a multi-mode fiber reduces speckle noise for low coherent light sources [8]. However, to achieve a speckle contrast of 1 %, a multi-mode fiber with a length between 10–30 m must be used. In this case, the intensity loss inside the fiber ranges around 20–30 % [28]. Though, since we use a laser with a coherence length of ≈ 300 m and a multi-mode fiber with a length of 2 m, only a multi-mode fiber is not sufficient to remove the speckle pattern from the image background. A detailed discussion about speckle reduction can be found in section 3A.

To illuminate the back aperture of the objective (FO), which focus the incoming light into the fiber, we walk the beam using two mirrors (M1, M2). We optimize the fiber alignment by monitoring the output power via a power meter (351 Hand-Held Power Meter, UDT Instruments (Gamma Scientific)) to ensure maximum coupling efficiency. Furthermore to achieve optimal coupling conditions, the numerical aperture (N.A.) of the fiber has to be equal or bigger than the N.A. of the focusing objective.

Second, we mount the collection objective (CO), the objective that collects scattered light from the diffuser, on a kinematic

mirror mount which is hold by a *xyz*-translation stage. Afterwards, this objective is placed into the beam path, with its front aperture facing the incoming light (details see section 6D, Fig. 6). By tilting the objective using the kinematic mirror mount and moving it with the *xyz*-translation stage, we ensure that the back reflection from the front aperture goes back to the light source. However, the back reflection should not hit the laser cavity to avoid strong intensity fluctuations. Furthermore, the correct position of the CO guarantees that the output beam is centered on the back aperture of the focusing objective (Fig. 1 b). Optimal alignment is achieved if the fiber output reaches its maximal value on the power meter.

In the third step, we introduce the stepper motor (SM) with the mounted ground glass diffuser (RGD) on its shaft in front of the collection objective. Again, back reflections from the RGD surface are centered on the laser. In this context, it must be highlighted, that the grit-polished side of the RGD must face away from the source. Subsequently, we set the distance between ground glass diffuser and collection objective to its working distance by using the translation stage of the motor (Table 1). This step is eminent to ensure parallel beam output on the back aperture of the collection objective.

Eventually, we place a 100 mm lens (L1) between mirror M2 and RGD to focus the light on the latter. This is necessary if the scattered light is collected with an objective or a single lens. We will discuss this issue in section 3B. After placing each optical component in the beam path, we fine tune the optical alignment by slightly changing the positions of the collection objective (in *xyz* and tilt) and of the fiber holder (in *xyz*) to place the focus of the focusing objective into the fiber entrance. We repeat this procedure until the maximum output power is obtained on the power meter.

B. Controlling the stepper motor using the Arduino Uno micro-controller and the motor driver module A4988

To operate the NEMA 17 stepper motor with 0.22 Nm torque, three components are needed: a DC motor power supply, an Arduino Uno micro-controller and a motor driver module A4988 (Fig. 2). In the first step, we connected the VMOT and the Ground pins of the driver module to a DC power supply, providing an voltage output of 10 V to run the motor. To protect the A4988 driver board from voltage spikes, we shunted a decoupling capacitor (capacity $C = 100 \mu\text{F}$) between the two power wires. The following four pins (2B, 2A, 1B, 1A) are used to connect the stepper motor with the A4988 driver card. For that purpose, we wired the pin couple 2A and 2B to coil 2 and couple 1A and 1B to coil 1 of the motor, respectively. The next two pins (VDD and Ground) are connected to the 5 V and GND pin on the Arduino Uno card, being the power supply for the A4988 driver card. The step pin (STEP) and direction pin (DIR), controlling the step size and rotation direction of the motor, respectively, are wired to Arduino card pins 3 and 4. These two pins are gated by the Arduino algorithm provided in section 6E. In the present setting, the stepper motor operates in full step mode. Eventually, we short-circuit the reset and sleep pin to avoid the stepper motor entering sleep mode. With the Arduino algorithm provided in section 6E the stepper motor operates continuously at its maximum rotation speed of 468 rpm, corresponding to a rotation frequency of 7.8 Hz. With an intrinsic motor resonance frequency of 0.5–1.5 Hz, the stepper motor runs smooth and stable at a rotation frequency of 7.8 Hz.

Fig. 2. Circuit diagram for stepper motor control using an Arduino Uno micro-controller and a motor driver module A4988.

C. Sample preparation for validation experiments and microscope details

To carry out validation experiments, we prepare a solution of polystyrene (PS) mono size particles of $(1.040 \pm 0.022) \mu\text{m}$ (mean $\pm$ SD), Lot No. 15879, Duke Scientific Corp., 4 % w/v) in deionized water or phosphate buffered saline (PBS, pH 7.4). Measurements are performed in a simple sandwich cell, consisting of two cover slips (lower cover slip: no. 1, Knittel Glass, 60 mm $\times$ 24 mm; upper cover slip: no. 1, Knittel Glass, 20 mm $\times$ 20 mm) separated by double sticky tape (Scotch, product no. 34-8509-3289-7, 3M). The cell volume is $\approx 10 \mu\text{L}$. We inject the diluted bead solution with a micro-pipette and seal the chamber by adding vacuum grease (Dow Corning) on both openings to prevent evaporation, which can result in a biased movement of the particles. To investigate motor-induced vibrations (section 3C) we incubate the sandwich cell overnight so PBS dissolved particles settle down and immobilize on the bottom cover slip. In contrast, measurements to measure the Brownian motion (section 3D) are performed immediately after injecting particles dissolved in deionized water into the sample cell.

Measurements are done using an Olympus IX70 inverted microscope, normally used for optical tweezers experiments, equipped with either an oil-immersion objective (Olympus UPlanFL 100x/1.30 Oil, $\infty/0.17$) or an free space objective (Olympus SLMPLN 50x/0.35, $\infty/0$) [29]. However, the 50x SLMPLN objective is only used to determine the highest camera frame rate at which speckle pattern reappear in the image background (section 3A). The prepared sample cell is mounted onto a piezo stage, which can be positioned in three dimensions over a range of 100 μm with nanometer accuracy using piezo actuators (P-561.3CD, Physik Instrumente). The sample is illuminated from above with parallel light produced from the optical fiber collimator (section 6A), and particles are imaged using a high-speed camera (MotionBLITZ EoSens Cube 7, Mikrotron) operating at 150 fps (section 3C) and 2000 fps (section 3D). We acquire images using the MotionBLITZDirector2 software and the resolution of the microscopy system is $(80 \pm 2) \text{ nm/pixel}$ (mean $\pm$ SD). The whole setup is built in a temperature controlled room at $(296 \pm 1) \text{ K}$ to ensure long-term stability and to reduce thermal drift effects [30]. For image analysis, we use the software UmUTracker [31].

3. DEMONSTRATION EXPERIMENTS

After aligning all optical components and controlling the rotation speed of the stepper motor with an Arduino card, we characterize the performance of our setup. First, we show that the rotating ground glass diffuser reduces the speckle pattern contrast providing high quality microscopic images. In this context, we propose a characterization procedure to determine the speckle contrast of the image background as a function of ground glass rotation speed and camera shutter time (section 3A). This procedure can be used to find the maximal camera frame rate for a given rotation speed with satisfying speckle contrast reduction. Second, we discuss experimental settings how to collect the scattered light from the RGG to maximize the optical fiber intensity output (section 3B). Third, we scrutinize if the rotating motor induces vibrations into the setup that might reduce the accuracy in vibration-sensitive measurements (section 3C). Eventually to feature the high laser light throughput, we record the Brownian motion of a freely diffusing particle using a 100x oil immersion objective at a camera frame rate of 2000 Hz and determine its diffusion constant (section 3D).

A. Setup calibration - Characterizing the speckle contrast in the image background

Reducing the speckle pattern using a RGG depends strongly on experimental settings such as illuminated area on the ground glass diffuser, its grid size on the polished side, its rotation speed but also on the camera shutter time and camera used to acquire images. If the laser beam is focused into the fiber entrance without ground glass diffuser, a characteristic speckle pattern can be observed (Fig. 3 a). If a stationary ground glass diffuser is introduced into the beam path the speckle field pattern still appears, but its properties depend now on the grid size of the ground glass diffuser (Fig. 3 b). By rotating the ground glass diffuser with a certain frequency, the speckle pattern disappears and an image without coherent noise and parasitic reflections can be recorded (Fig. 3 c). Detailed descriptions about how the mentioned experimental parameters affect the coherence length of the laser and therefore the speckle pattern can be found elsewhere [12, 13, 16].

Next, we discuss the efficiency of speckle reduction for our setup. In general, the speckle reduction factor equals $1/\sqrt{N}$, with N being the number of recorded uncorrelated speckle patterns with equal mean intensity within the camera shutter time t_{Shutter} [27, 28]. In numbers this means, if 10 000 speckle patterns are recorded, the speckle contrast in the image background is reduced by a factor of 100. However, calculating N to achieve sufficient speckle reduction is challenging, since it depends on many system parameters like grid size and rotation speed of the ground glass, camera shutter time, camera chip but also on the imaging system of the microscope and will be different for each setup. To overcome this inconvenience we propose a characterization procedure to measure the normalized speckle contrast C_{norm} in dependence of the rotation speed of the ground glass at a fixed camera frame rate. The normalized speckle contrast is defined as [28, 32]:

Fig. 3. Microscopic images acquired under different experimental conditions. (a) Speckle pattern produced by laser light without rotating ground glass diffuser. (b) Speckle pattern produced by laser light which is scattered by a stationary ground glass diffuser. (c) Image with rotating ground glass diffuser. Since the RGG reduce the spatial coherence of the laser light, no speckle pattern is visible. (d) Speckle contrast vs. ground glass rotation speed to determine the critical rotation speed v_c above which speckle pattern are removed from the image background. (e) Image sequence acquired for a rotation speed above ($v_{\text{RGG}} > v_c$) and below ($v_{\text{RGG}} < v_c$) the critical rotation speed. The partial reappearance of speckle pattern in case $v_{\text{RGG}} < v_c$ can be explained by intrinsic properties of the stepper motor.

$$C_{\text{norm}}(v_{\text{RGG}}) = \frac{C(v_{\text{RGG}}) - C(v_{\text{RGG}} = \max.)}{C(v_{\text{RGG}} = 0)} \quad (1)$$

$$C(v_{\text{RGG}}) = \frac{1}{I_{\text{mean}}} \sqrt{\frac{1}{M} \sum_{i=1}^M (I_i - I_{\text{mean}})^2} \quad (2)$$

$$I_{\text{mean}} = \frac{1}{M} \sum_{i=1}^M I_i, \quad (3)$$

where $C(v_{\text{RGG}})$ is the mean speckle contrast from an image sequence acquired at a given rotation speed v_{RGG} , $C(v_{\text{RGG}} = \max.)$ is the speckle contrast at $v_{\text{RGG}} = \max.$, $C(v_{\text{RGG}} = 0)$ is the speckle contrast at $v_{\text{RGG}} = 0$ (this corresponds to the contrast of a fully developed speckle pattern), I_{mean} is the mean image intensity of M acquired images at a certain v_{RGG} and I_i is the image intensity of the i -th image at v_{RGG} .

In this context, we will now summarize and discuss experimental parameters used for the presented setup to measure the normalized speckle contrast which depends on v_{RGG} and t_{Shutter} . By considering the surface properties of the ground glass diffuser, the most efficient way to reduce speckle contrast is to use the smallest grid size available [12]. We choose a grid size of 1500. The illuminated area on the RGG has a diameter $d_{\text{ill. area}} \approx 60 \mu\text{m}$ (Table 1) and the laser beam hits the RGG at

22.4 mm above the rotation center. To measure the normalized speckle contrast we vary the rotation speed of the RGG in the range of 0.014–1.1 m/s (corresponding to 6–468 rpm) for camera shutter times of [1000, 500, 100, 64] μs (Fig. 3 d). Furthermore, for each shutter time we record the speckle contrast for the stationary RGG. For data representation we do not show the speckle contrast of a stationary ground glass ($C(v_{\text{RGG}} = 0) = 1$). With decreasing rotation speed, the speckle contrast increases since less independent speckle patterns are acquired within the camera shutter time. However, for a camera shutter time of 1000 μs (green data points) and 500 μs (blue data points) we measure a speckle contrast less than 4 % and 8 % at the lowest rotation speeds ($v_{\text{RGG}} = 0.014$ m/s), respectively. If the camera shutter time is set to 100 μs (black data points) an 8 % increase in speckle contrast is already reached at a rotation speed of 0.16 m/s. By reducing the camera shutter time even further to 64 μs (cyan data points), we find a 8 % speckle contrast increase at $v_{\text{RGG}} = 0.28$ m/s. This indicates that at low camera shutter times the effective speckle reduction is limited by the maximum rotation speed of the stepper motor rotating the ground glass diffuser.

Additionally, we observe a significant increase in standard deviation of the normalized speckle contrast if the RGG is slower than a certain critical rotation speed (Fig. 3 e, v_c , red dashed line). This can be explained by intrinsic properties of the stepper motor. To rotate the ground glass diffuser 360° our stepper motor executes 200 discrete steps with a step size of 1.8° (Fig. 3 e, bottom). However, the rotation speed between 2 steps is not constant since the motor slows down after traveling 1.8° as it is "waiting" for the next signal. If the motor moves slower than v_c the RGG becomes for a short period of time stationary and an image with an almost fully developed speckle pattern is recorded (Fig. 3 e, middle). These images possess a higher speckle contrast and give rise to the increased standard deviation. In contrast, if $v_{\text{RGG}} \geq v_c$ the motor is moving fast even before a new signal is sent, resulting in a speckle free image sequence (Fig. 3 e, top).

Based on our characterization method, we advise users to operate their stepper motor at maximum rotation speed to avoid artifacts from the step-like motion of the motor and from an increase speckle contrast at short camera shutter times. In the presented setup, the stepper motor operates at 468 rpm, corresponding to an angular velocity 1.1 m/s. With these settings and a maximal intensity output of 22 mW, we can record speckle free images up to 10 000 kHz. In case of shorter shutter times (e.g. 64 μs), speckle contrast rises significantly, necessitating faster rotation speeds. In this case, we suggest the usage of a stepper motor with higher torque or a DC motor. Our Matlab program to calculate the speckle contrast $C(v_{\text{RGG}})$ from camera images can be downloaded from Ref. [33].

B. Choice of collection objective and output intensity

Besides removing speckle contrast effectively, a high fiber intensity output to guaranty enough light for measurements at high camera frame rates is desirable. As the RGG scatters incoming light significantly, the closer to the scattering surface the light is collected, the higher the collection efficiency is. Therefore, we use an ordinary microscope objective instead of normal lenses since small distances between RGG and objective can be realized. The only requirement for the used collection objective (CO, Fig. 1 b) is a large front aperture to ensure effective light collection. For that reason, we choose a standard objective (Olympus Plan N 10x). To judge its collection efficiency, we determine the beam

Table 1. Comparison of fiber intensity output for two light collecting options: objective and single lens (WD = working distance, N.A. = numerical aperture, d_{Spot} = theoretical focal spot diameter, $d_{\text{ill. area}}$ = diameter of illuminated area on RGG, P_x = laser intensity).

Objective	Olympus Plan N 10x	single lens
WD	10.6 mm	60 mm
N.A.	0.25	0.21
d_{Spot}	3.0 μm	35 μm
$d_{\text{ill. area}}$	60 μm	60 μm
$P_{\text{before RGG}}$	10.8 mW	10.8 mW
$P_{\text{fiber output}}$	5.2 mW	3.1 mW
coupling eff.	$\approx 48 \%$	$\approx 29 \%$

power to be 10.8 mW in front of the RGG and measure the fiber output power using a power meter, revealing a fiber output power of 5.2 mW (Table 1). This efficient light throughput of 48 % enables us to perform imaging up to 10 000 Hz at only tens of MilliWatts laser power, preventing the setup and sample from laser-induced damaged.

To achieve optimal light throughput we ensure that the output beam from the collection objective is parallel since this beam will be guided into the optical fiber by the focusing objective. However, objectives produce only collimated light under certain conditions. We will clarify this in the following. Consider a laser beam with a Gaussian intensity distribution entering the back aperture of an objective. If the beam diameter equals the objectives entrance pupil diameter, the light will be focused at the working distance of the objective. The spot size can then be approximated by the diffraction limit: $d_{\text{Spot}} \approx 1.22\lambda / \text{N.A.}$ which is $\approx 3 \mu\text{m}$ for the 10x collection objective (Table 1). Vice versa, an objective produces a collimated beam if the diameter of the imaged light source is less than or equal to d_{Spot} . In other words, if the diameter of the illuminated area ($d_{\text{ill. area}}$) on the RGG is bigger than d_{Spot} , the output beam from the collection objective will be divergent. The divergence becomes more pronounced the more the ratio between $d_{\text{ill. area}}$ and d_{Spot} differs from unity. The same condition applies if a lens is used for light collection (see next paragraph and Table 1). To meet the discussed requirement for collimated beam output, we introduced lens L1 into the light path to reduce the diameter of the illuminated area on the RGG from 1300 μm to 60 μm . In this case, the collection objective produces slightly divergent light. To compensate the divergence, we place the collection objective marginally closer to the RGG than its working distance. Despite focusing the laser beam on the RGG, the illumination area is still big enough to realize effective speckle averaging for camera frame rates up to 10 000 Hz.

To highlight the high light collection efficiency by using only one objective, we compare this scenario to a single lens collector as reported in literature [16–20]. For that purpose, we place a 60 mm lens (LA1134-A; Thorlabs) behind the ground glass diffuser, possessing a similar numerical aperture (N.A.) as the used objective (Table 1). Again, the laser intensity in front of the RGG is set to 10.8 mW. The lens output is a slightly divergent beam (see discussion above) with a diameter of 25.4 mm and an intensity of 4.4 mW. Guiding this beam into the fiber produces

an output of only 0.2 mW as the majority of the light is truncated by the back aperture of the focusing objective. Therefore, we decrease the beam size using a negative beam magnification (LA1509-A, LA1134-A; Thorlabs), revealing an fiber output of 3.1 mW, which is $\approx 40\%$ lower than the achieved intensity using an objective for light collection. Furthermore, using the same objective for light collection and light focusing into the fiber has the advantage, that no further optics are needed to manipulate the beam diameter since the collimated output beam has the same diameter as the back aperture of the focusing objective.

C. Vibrations caused by the RGG do not impair image quality

In case the gear and the ground glass diffuser are not absolutely centered to each other (section 6D), the RGG can induce vibrations that might perturb sensitive measurements. This problem becomes pronounced if the stepper motor is placed on the same optical table as the measurement microscope. To estimate the impact of vibrations on measurements when the stepper motor is running at its maximum rotation speed, we immobilize 1 μm PS particles on the bottom cover slip of the sandwich chamber (section 2C) and image their position using laser (with RGG) and LED (without RGG) illumination at 150 Hz (Fig. 4 a). To compare the position width using laser (blue data points) and LED illumination (black data points), we analyze their profiles along the x and y direction using a Gaussian fit (Fig. 4 b-e). As a result we obtain a mean full width at half maximum (FWHM) of (31.5 ± 1.5) nm in x and (27 ± 2) nm in y direction, respectively. By considering a camera pixel size of 80 nm/pixel, the bead position varies within one pixel and no difference in the particle center position using either laser or LED illumination can be found. Consequently, rotation induced vibrations at a rotation speed of 468 rpm can be neglected.

Eventually, it should be mentioned that stepper motors possess a resonance frequency at which they produce heavy vibrations. For example, the motor used in this setup has its resonance frequency around 0.5–1.5 Hz. At this rotation speed, vibrations will propagate to the sample and can be measured. Therefore, if the stepper motor is operated around its resonance frequency, we recommend to build the laser illumination setup with the RGG on a separate optical table, since longer distances can convincingly be bridged using a longer optical fiber.

D. Brownian motion analysis

High speed imaging requires short shutter times, which in turn demands an adequate amount of light reaching the sample under study. To highlight the potential of our speckle-free laser illumination setup, we track freely diffusing 1 μm PS particles in deionized water in two dimensions (Fig. 5 a). To achieve optimal spatial resolution, we use a 100x oil immersion objective and acquire images at a camera frame rate of 2000 Hz (camera shutter time 500 μs). The laser light output at the fiber collimator is set to 22 mW. From the trajectories of four independent diffusing particles, we calculate the 2D mean square displacement (MSD, Fig. 5 b, black data) and fit the data with a linear function (blue line). By considering the relation between MSD and diffusion constant D : $\text{MSD}(\Delta t) = 4D\Delta t$, the slope of the linear fit equals $4D$ revealing a diffusion constant of $D = (0.45 \pm 0.05) \mu\text{m}^2/\text{s}$. This value can be theoretically validated by using the Einstein-Stokes relation [34]: $D = k_B T \cdot (3\pi\eta d)^{-1}$, with $k_B T$ being the thermal energy, η the viscosity of water and d the particle diameter. With a temperature $T = 296$ K, a viscosity $\eta = 0.932 \times 10^{-3}$ Pa s and a particle diameter $d = 1.04 \mu\text{m}$, we calculate the theoretical

Fig. 4. Influence of vibrations caused by the RGG when determining the particle center position. (a) Comparison of particle center positions estimated under LED (without RGG, black data) and laser (with RGG, blue data) illumination acquired at a frame rate of 150 Hz (camera shutter time 6.7 ms). (b-c) Quantitative analysis of particle movement in the x and y direction under LED illumination (black data points). A Gaussian function (blue line) is fitted to the data, revealing a full width at half maximum of 33 nm (x direction, $R^2 = 0.97$) and 30 nm (y direction, $R^2 = 0.98$), respectively. (d-e) Position analysis of the same particle as in (b-c) but with the RGG and laser illumination (black data points). The Gaussian fit (blue line) produces a full width at half maximum of 29 nm (x direction, $R^2 = 0.96$) and 25 nm (y direction, $R^2 = 0.98$).

diffusion constant to $D_{\text{theo}} = 0.45 \mu\text{m}^2/\text{s}$, showing agreement with our experimental value.

Fig. 5. (a) Typical 2D trajectory of a PS particle with a diameter of 1 μm freely diffusing in deionized water. Data acquired at a frame rate of 2000 Hz resulting in 23000 data points. (b) 2D mean square displacement (black data points) versus time. To calculate the MSD four independent particle trajectories are used. By fitting the MSD with a linear function (blue line, $R^2 = 0.94$) the diffusion constant is assessed to $D = (0.45 \pm 0.05) \mu\text{m}^2/\text{s}$.

4. CONCLUSION

We presented a step-by-step guide to build a speckle-free laser illumination setup using a rotating glass diffuser mounted on a stepper motor, that is easy to realize, compact and robust. Furthermore, by using only one objective to collect the scattered light behind the ground glass diffuser, we improved the light throughput by $\approx 40\%$ compared to using single lenses collector. Costs for the stepper motor, its controlling unit and the objective are below 500 USD being significantly cheaper than

commercial solutions. We provided detailed experimental information for setup construction, enabling the user to record speckle-free images up to 10 000 Hz. We characterized our setup in terms of image quality, speckle contrast, light throughput and motor-induced vibrations. To highlight a light throughput of $\approx 48\%$, we recorded the Brownian motion of micro-particles at a frame rate of 2000 Hz using a 100x oil immersion objective and determined the diffusion constant. By providing microscopic images with highest quality at high camera frame rates and moderate laser intensities, our setup can be used for example to investigate surface-attached bacteria under physiological flow conditions using digital holography without applying image post-processing such as background subtraction or filtering.

5. FUNDING INFORMATION

T.S. acknowledges financial support from the German Research Foundation (DFG) via a postdoctoral fellowship. This work was supported by the Swedish Research Council (2013-5379) and from the Kempe foundation to M.A.

6. APPENDIX: COMPONENT LISTS, INSTALLATION DETAILS AND ARDUINO ALGORITHM

A. Optical Components

1. Cobolt Calypso™50 laser ($\lambda = 491$ nm, max. output power $P = 45$ mW), Cobolt AB, Solona, Sweden – Fig. 1: Laser
2. Optical beam shutter, Thorlabs (SH05) – Fig. 1: Shutter
3. Kinematic mirror mount for $\varnothing 1''$ optics, Thorlabs (KM100) – Fig. 1: M1, M2, L1 & Fig. 6: Mount
4. 30 mm cage plate for $\varnothing 1''$ optics, Thorlabs (CP06)
5. $\varnothing 1''$ broad band dielectric mirror, 400–750 nm, Thorlabs (BB1-E02) – Fig. 1
6. $\varnothing 1''$ N-BK7 plano-convex lens (AR coating: 350–700 nm), $f = 100$ mm, Thorlabs (LA1509-A) – Fig. 1: L1
7. $\varnothing 2''$ unmounted N-BK7 ground glass diffuser, 1500 grit, Thorlabs (DG20-1500) – Fig. 1 & Fig. 6: RGG
8. 10x Olympus plan achromat objective, 0.25 NA, 10.6 mm WD, Thorlabs (RMS10X), – Fig. 1 & Fig. 6: CO and FO
9. Fiber launch with FC-connectorized fiber holder, Thorlabs (MBT613D)
10. $\varnothing 400$ μ m, 0.39 NA, FC/PC to SMA multimode hybrid fibre optic patch cable, Thorlabs (M76L02) – Fig. 1 (OF)
11. Aspheric FiberPort, SMA, $f = 7.5$ mm, 350–700 nm, $\varnothing 1.23$ mm waist, Thorlabs (PAF-SMA-7-A)

B. Mechanical Components

- Triple divide xyz translation stage, 28 mm travel, Newport (9064-XYZ-M) – Fig. 1 & Fig. 6: TS
- NEMA 17 high torque stepper motor, 42H33H-1334A, 0.22 N m torque, Amazon – Fig. 1 & Fig. 6: SM
- Aluminium V-belt pulley Reely bore diameter: 5 mm, outer diameter: 20 mm, Conrad (238325-62) – Fig. 1 & Fig. 6: Gear

- Switchable, heavy-duty magnetic base, Thorlabs (MB175/M) – Fig. 1 & Fig. 6: Base
- 3D printed motor holder, CAD file and mechanical drawing can be downloaded from Ref. [33] – Fig. 1 & Fig. 6: MH
- 3D printed holder for $\varnothing 1/2''$ stainless steel optical posts, CAD file and mechanical drawing can be downloaded from Ref. [33] – Fig. 1 & Fig. 6: PH
- 3D printed holder for RMS threaded objectives, CAD file and mechanical drawing can be downloaded from Ref. [33] – Fig. 1 & Fig. 6: OR

C. Electrical Components

- GW Instek GPS-30300 DC power supply, Gw Instek
- Arduino Uno R3 micro-controller, Amazon – Fig. 2
- Stepstick stepper motor driver module A4988, Amazon – Fig. 2
- Breadboard, K&H MFG. Co., LTD (RH-21B) – Fig. 2
- Panasonic 100 μ F radial electrolyte capacitor, Conrad (421945-62) – Fig. 2

D. Installation details for stepper motor with rotating ground diffuser and collection objective

In this section, we provide step-by-step information how to install the stepper motor (SM) on the translation stage and how to glue the ground glass diffuser (RGG) to a gear, which will be mounted on the D-cut motor shaft. Furthermore, we explain how the collection objective (CO) is installed on the xyz -translations stage (Fig. 6).

We start with the assembly of stepper motor and ground glass diffuser. In the first step, a single translation stage is mounted on an optical table (Fig. 6 a). Second, the lower surface of the 3D printed motor holder is screwed to the stage using four screws (Fig. 6 b). The CAD file and the mechanical drawing file of the holder can be downloaded from Ref. [33]. Third, the stepper motor (SM) is mounted to the front side of the holder (Fig. 6 c). Before tighten the gear on the D-cut motor shaft with a set screw (Fig. 6 d), we glue the ground glass diffuser with its smooth surface to the gear. For this we use a cyanoacrylate-based adhesive. The gluing step is crucial and must be carried out with highest precision to ensure that gear and ground glass diffuser are centered. If the centers are misaligned, the motor will be imbalanced and vibrations at high rotation speeds occur.

To control the position and orientation of the collection objective (CO) using a xyz -translation stage, the latter is pre-assembled on a magnetic base containing a switchable magnet. Afterwards, the 3D printed holder for $\varnothing 1/2''$ stainless steel optical posts and the post are mounted on the z -translation stage (Fig. 6 e). In the next step, we mount a kinematic mirror mount (Mount) on the tip of the post (Fig. 6 f) to control the tilt of the collection objective. Thereafter, we pre-assemble the collection objective (CO) and the 3D printed holder for RMS threaded objectives (OR) (Fig. 6 g, Inset). In the final step, we insert the objective holder into the kinematic mirror mount and tighten it with a set screw (Fig. 6 g).

Fig. 6. Step-by-step manual to assemble the stepper motor (SM) with the ground glass diffuser (RGG) and the collection objective (CO) with its holder (OR).

E. Arduino Algorithm

Here, we provide the Arduino algorithm 1 to control the rotation speed of the stepper motor (NEMA 17, section 2B), written in Arduino Software 1.8.1. With the provided settings, the motor runs at 468 rpm. However, to achieve this maximum rotation speed, the stepper motor needs to accelerate slowly by starting at a low speed. This is caused by the motor inertia and the mass of the ground glass diffuser. The first 3 lines of the code define motor-dependent input values, such as initial time for one revolution ("*const int InitialTimeStep = 700*"), acceleration time step ("*const int AcceleratingTimeStep = 2*;) and time per revolution after the final acceleration step ("*const int FinalTimeStep = 128*;" in [ms]). The latter sets the maximum rotation speed. Afterwards, the step ("*stepPin*") and direction ("*dirPin*") pin on the Arduino board are specified and set as output. The loop starting in line 13 controls the acceleration process based on the three predefined input values mentioned above. Once the acceleration has reached the "*FinalTimeStep*", the motor will run with this value for 1000 min (line 24). The algorithm code can be downloaded from Ref. [33].

Algorithm 1. Arduino algorithm to control the rotation speed of a stepper motor.

```

1: const int InitialTimeStep =700;
2: const int AcceleratingTimeStep= 2;
3: const int FinalTimeStep= 128;
4:
5: const int stepPin = 3;           ▷ defines pin numbers
6: const int dirPin = 4;
7:
8: void setup() {                  ▷ sets the two pre-defined pins as output
9:   pinMode(stepPin, OUTPUT);
10:  pinMode(dirPin, OUTPUT);
11: }
12:
13: void loop() {                  ▷ Acceleration part
14:   digitalWrite(dirPin, HIGH);
15:   for (int i = InitialTimeStep; i > FinalTimeStep; i = i - AcceleratingTimeStep){}
16:   for (int x = 0; x < 200; x++) {
17:     digitalWrite(stepPin, HIGH);
18:     delayMicroseconds(i);
19:     digitalWrite(stepPin, LOW);
20:     delayMicroseconds(i);
21:   }
22: }
23:
24: for (int i = InitialTimeStep; i < 3600000; i = i++) { ▷ Defines run time of script
25:   for (int x = 0; x < 200; x++) {
26:     digitalWrite(stepPin, HIGH);
27:     delayMicroseconds(FinalTimeStep);
28:     digitalWrite(stepPin, LOW);
29:     delayMicroseconds(FinalTimeStep);
30:   }
31: }
32: }

```

REFERENCES

1. B. Redding, M. a. Choma, and H. Cao, "Speckle-free laser imaging using random laser illumination," *Nat. Photonics* **6**, 355–359 (2012).
2. G. A. Alphonse and D. B. Gilbert, "Super-luminescent diode," (1989).
3. L. Goldberg and D. Mehuys, "High power superluminescent diode source," *Electronic Letters* **30**, 1682–1684 (1994).
4. D. K. Smith and J. A. Casey, "Laser-driven light source," (2011).
5. Dainty J. C., "Laser speckle and related phenomena," *Topics in Applied Physics* **9**, 298 (1975).

6. J. W. Goodman, "Some fundamental properties of speckle," *Journal of the Optical Society of America* **66**, 1145–1150 (1976).
7. J. W. Goodman, *Speckle phenomena in optics: theory and applications* (2005).
8. E. Gratton and M. J. VandeVen, "Laser Sources for Confocal Microscopy," *Handbook of Biological Confocal Microscopy* pp. 80–125 (2006).
9. G. W. Ellis, "A fiber-optic phase-randomizer for microscope illumination by laser," *Journal of Cell Biology* **83**, 303a (1979).
10. C. M. Davenport and A. F. Gmitro, "Angioscopic fluorescence imaging system," (1992), vol. 1649, pp. 192–202.
11. S. Bains, "Holographic optics: for when less is more," *Laser focus world* **29**, 151–156 (1993).
12. T. Asakura, "Spatial coherence of laser light passed through rotating ground glass," *Opto-electronics* **2**, 115–123 (1970).
13. L. E. Estes, L. M. Narducci, and R. A. Tuft, "Scattering of Light from a Rotating Ground Glass," *Journal of the Optical Society of America* **61**, 1301–1306 (1971).
14. A. Kumar, J. Banerji, and R. P. Singh, "Intensity correlation properties of high-order optical vortices passing through a rotating ground-glass plate," *Optics letters* **35**, 3841–3 (2010).
15. Y. H. Zhai, X. H. Chen, D. Zhang, and L. A. Wu, "Two-photon interference with true thermal light," *Physical Review A - Atomic, Molecular, and Optical Physics* **72** (2005).
16. F. Dubois, M.-L. Novella Requena, C. Minetti, O. Monnom, and E. Istasse, "Partial spatial coherence effects in digital holographic microscopy with a laser source," *Appl. Opt.* **43**, 1131–1139 (2004).
17. A. Fougères and W. Noh, "Measurement of phase differences between two partially coherent coherent fields," *Physical Review A - Atomic, Molecular, and Optical Physics* **49**, 530–535 (1994).
18. M. A. A. Neil, A. Squire, R. Juškaitis, P. I. H. Bastiaens, and T. Wilson, "Wide-field optically sectioning fluorescence microscopy with laser illumination," *Journal of Microscopy* **197**, 1–4 (2000).
19. G. Scarcelli, A. Valencia, and Y. Shih, "Experimental study of the momentum correlation of a pseudothermal field in the photon-counting regime," *Physical Review A - Atomic, Molecular, and Optical Physics* **70**, 1–4 (2004).
20. W. Wang, Z. Duan, S. G. Hanson, Y. Miyamoto, and M. Takeda, "Experimental study of coherence vortices: Local properties of phase singularities in a spatial coherence function," *Physical Review Letters* **96**, 1–4 (2006).
21. J. H. Shapiro, "Computational ghost imaging," *Physical Review A - Atomic, Molecular, and Optical Physics* **78**, 1–4 (2008).
22. F. Ferri, D. Magatti, A. Gatti, M. Bache, E. Brambilla, and L. A. Lugiato, "High-resolution ghost image and ghost diffraction experiments with thermal light," *Physical Review Letters* **94**, 2–5 (2005).
23. M. Zhang, Q. Wei, X. Shen, Y. Liu, H. Liu, J. Cheng, and S. Han, "Lensless Fourier-transform ghost imaging with classical incoherent light," *Physical Review A - Atomic, Molecular, and Optical Physics* **75**, 1–4 (2007).
24. T. Funatsu, Y. Harada, M. Tokunaga, K. Saito, and T. Yanagida, "Imaging of single fluorescent molecules and individual {ATP} turnovers by single myosin molecules in aqueous solution." (1995).
25. M. K. Kim, *Digital Holographic Microscopy* (Springer New York, New York, NY, 2011), pp. 149–190.
26. D. Huang, E. A. Swanson, C. P. Lin, J. S. Schuman, W. G. Stinson, W. Chang, M. R. Hee, T. Flotte, K. Gregory, C. A. Puliafito, and al. Et, "Optical coherence tomography," *Science* **254**, 1178 LP – 1181 (1991).
27. J. I. Trisnadi, "Speckle contrast reduction in laser projection displays," *Proceedings of the Society of Photo-Optical Instrumentation Engineers (Spie)* **4657**, 131–137 (2002).
28. J. G. Manni and J. W. Goodman, "Versatile method for achieving 1% speckle contrast in large-venue laser projection displays using a stationary multimode optical fiber," *Optics Express* **20**, 11288–11315 (2012).
29. E. Fällman, S. Schedin, J. Jass, M. Andersson, B. E. Uhlin, and O. Axner, "Optical tweezers based force measurement system for quantitating binding interactions: system design and application for the study of bacterial adhesion," *Biosensors and Bioelectronics* **19**, 1429–1437 (2004).
30. M. Andersson, F. Czerwinski, and L. B. Oddershede, "Optimizing active and passive calibration of optical tweezers," *Journal of Optics* **13**, 044020 (2011).
31. H. Zhang, T. Stangner, K. Wiklund, A. Rodriguez, and M. Andersson, "UmUTracker: A versatile MATLAB program for automated particle tracking of 2D light microscopy or 3D digital holography data," (2017).
32. D. S. Mehta, D. N. Naik, R. K. Singh, and M. Takeda, "Laser speckle reduction by multimode optical fiber bundle with combined temporal, spatial, and angular diversity," *Applied Optics* **51**, 1894 (2012).
33. T. Stangner, "Supplementary Information," <https://doi.org/10.6084/m9.figshare.4758451.v2> (2017). [Retrieved: 30 May 2017].
34. A. Einstein, "Über die von der molekularkinetischen Theorie der Wärme geforderte Bewegung von in ruhenden Flüssigkeiten suspendierten Teilchen," *Annalen der Physik* **322**, 549–560 (1905).